

Single beam spectrophotometer is inconvenient because The sample and blank must be placed alternately in the light path. For measurements at multiple wavelengths, the blank must be run at each .wavelength