

In fact, the spectrum produced is an absorbance spectrum rather than a transmittance spectrum, but it .can be readily converted into a transmittance spectrum if the instrument is attached to a computer