

Temperature Control and Data Processing The temperature recorded in a thermogram is ideally the actual temperature be obtained by immersing a small thermocouple directly in the sample. Due to some possible catalytic decomposition of samples, potential contamination of samples, and mass measurement errors resulting from the thermocouple leads. Modern TGA systems use a computerized temperature control routine that automatically compares the voltage.