

The standard addition technique is used for non-repetitive analyses or to correct for matrix effects that cannot be accommodated in the calibration standards. It is then extrapolated to the concentration axis to acquire the concentration of sample solution. In this method a known amount of the standard solution of known increasing concentrations of the analyte is added to a number of aliquots of the sample solution.